



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

DC SCIENTIFIC, INC.
6750 McLean Way, Suite D
Glen Burnie MD 21060
Doug Sisson Phone: 410 863 1700

CALIBRATION

Valid To: February 28, 2020

Certificate Number: 4364.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations¹:

I. Fluid Quantities

Parameter/Equipment	Range	CMC ^{2,3} (±)	Comments
Viscosity Working Viscometers Direct Flow – Suspended Level (Transparent Liquids)	<10 mm ² /s (10 to 100) mm ² /s (100 to 1000) mm ² /s (1000 to 10 000) mm ² /s (10 000 to 100 000) mm ² /s	0.36 % 0.37 % 0.38 % 0.40 % 0.42 %	ASTM D446 ⁴ ISO 3104 ⁴
Viscosity Working Viscometers Reverse Flow – (Transparent and Opaque Liquids)	<10 mm ² /s (10 to 100) mm ² /s (100 to 1000) mm ² /s (1000 to 10 000) mm ² /s (10 000 to 100 000) mm ² /s	0.36 % 0.37 % 0.38 % 0.40 % 0.42 %	ASTM D446 ⁴ ISO 3104 ⁴
Viscosity Working Viscometers Modified Ostwald	<10 mm ² /s (10 to 100) mm ² /s (100 to 1000) mm ² /s (1000 to 10 000) mm ² /s (10 000 to 100 000) mm ² /s	0.36 % 0.37 % 0.38 % 0.40 % 0.42 %	ASTM D446 ⁴ ISO 3104 ⁴

¹ This laboratory offers commercial calibration services.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ CMCs expressed in percentages of reading.

⁴ Viscometers are calibrated to ASTM D446/ISO 3104 using viscosity standards to ASTM D2162.



Accredited Laboratory

A2LA has accredited

DC SCIENTIFIC, INC.

Glen Burnie, MD

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets *R205 – Specific Requirements: Calibration Laboratory Accreditation Program*. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009).



Presented this 20th day of November 2017.

A handwritten signature in black ink, written over a horizontal line.

President and CEO
For the Accreditation Council
Certificate Number 4364.01
Valid to February 28, 2020

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.